


# Rabbit monoclonal antibody against PMS2 (Clone ISR019)



**Figure 1** Colorectal cancer stained with anti-PMS2 antibody.

#### Product identification

ISC116-R3	3 ml ready-to-use (RTU)
ISC116-R7	7 ml ready-to-use (RTU)
ISC116-C02	0.2 ml concentrated
ISC116-C05	0.5 ml concentrated

# Summary and explanation

PMS2 is the component of the post-replicative DNA mismatch repair system (MMR). Except PMS2, MMR gene include MLH1, MSH2 and MSH6. PMS2 is located on chromosome 7 and spans~38000 base pairs. The full-length PMS2 protein product is translated from 15 exons and is 862 amino acids in length. PMS2 forms a heterodimer complex with MLH1 via a C-terminal interaction domain. This interaction is absolutely obligatory for the protein stability of PMS2 in vivo, and its dysfunction leads to degradation or loss of PMS2. Compared with other MMR germline mutations, PMS2 germline mutation has the feature of later onset, weaker family history, lower risk for cancer and hard for detection.

# Principle of the procedure

The stated primary antibody is suitable for immunohistochemical staining of FFPE tissue sections based on specific antigen-antibody reaction. Using a detection system linked to HRP or alkaline phosphatase the antigen visualization is performed via specific binding of the primary antibody. Secondary antibody is binding to the primary antibody, and the enzyme complex labels this complex. The enzymatic activation of the chromogen results in a visible reaction product at the antigen site. Each step is incubated for a precise time and temperature and requires interposed washing steps. The specimen may then be counterstained. Results are interpreted using a light microscope.

Materials provided
--------------------

Primary antibody	PMS2
Host	Rabbit
Subclass	lgG
Clone	ISR019
Species Cross-reactivity	Human. Others-not
	known
Applications	Immunohistochemistry
Epitope Retrieval	Heat-induced epitope
	retrieval
Ready-to-use antibody	Prediluted antibody in
	antibody diluent buffer
Recommended working	1:50 to 1:100
dilution range	

Product label shows the specific lot number.

Prediluted antibody is ready-to-use and optimized for staining. No further dilution, reconstitution, mixing, or titration is needed.

Antibody concentrate is optimized for dilution within dilution range using InSituChem® Antibody Diluent for IHC (Cat. No. D005-50). Indicated dilution range should be considered as recommendation and depends on different facts (tissue, fixation, incubation conditions, etc.). Optimum dilution to be determined in user's own system.

#### Materials required but not provided

The following materials may be required for staining but are not provided with the primary antibody.

- Positive and negative controls
- Microscope slides (positively charged) and cover slips
- Water bath
- Humidified chamber
- Staining jars
- Stopwatch
- Xylene or xylene substitute
- Ethanol
- Deionized or distilled water
- Antigen retrieval reagent, e.g. Antigen Enhancer (HIER buffer), Cat. No. D004-500
- Detection system, e.g. IHC Complete Detection system (Goat anti mouse/rabbit HRP, DAB staining), Cat. No. D001-15
- Wash buffer: e.g. IHC Wash Buffer, Cat. No. D003-500
- Tap water/bluing reagent (e.g. ammonia water)
- Light microscope

# Storage and handling

#### Store at 2 - 8 °C.

When stored correctly antibody is stable to the expiration date indicated on the vial. Do not use after expiration date.

PMS2 En Rev. 1 IM Page 1 of 2



To ensure proper reagent delivery and stability of the antibody, replace the dispenser cap after every use and immediately place the bottle cool in an upright position.

# Staining procedure

- 1. Cut 3-4 µm section of formalin-fixed paraffinembedded tissue and place on positively charged
- 2. Dry at 65°C for 2 hours.
- 3. Deparaffinize, rehydrate, and epitope retrieve. Upon completion, rinse with 3 changes of distilled or deionized water.
- 4. If using HRP detection system, place slides in peroxide block for 10 minutes; rinse. If using AP detection system, omit this step.
- 5. Apply the antibody and incubate for 60 minutes; rinse.
- 6. Apply the InSituVison™ Polymer Rabbit/Mouse Detection System for 30 minutes; rinse.
- 7. Apply ample amount of DAB or AEC chromogen and incubate; rinse.
- 8. Dehydrate and coverslip.

# Interpretation of results

The immunostaining procedure causes a colored reaction product to precipitate at the antigen sites localized by the primary antibody.

#### Cellular localization: Cell nucleus.

Α qualified pathologist experienced in immunohistochemistry procedures must evaluate positive and negative tissue controls before interpreting patient specimens.

Positive staining intensity should be assessed within the context of any background staining of the negative reagent control.

Note: A negative result means that the antigen in guestion was not detected, but not that the antigen is not present in the cells/tissues tested. An antibody panel may be used to support the results in some circumstances. Additionally, the morphology of each tissue sample should be examined utilizing a hematoxylin and eosin stained section. A qualified pathologist must interpret the patient's morphologic findings and pertinent clinical data.

# Warnings and precautions

- 1. Application only by qualified and trained personnel.
- 2. There are no estimated health risks, if the product is used as directed. MSDS is available on request.
- 3. Product contains sodium azide as preservative. Pure sodium azide is toxic. The concentration of sodium azide in this reagent is < 0.1 % and is not classified hazardous. See MSDS.
- 4. As with any product derived from biological sources, proper handling procedures should be used.
- 5. Do not use reagents after expiration date.

- 6. Take reasonable precautions when handling reagents. Use protective clothing and gloves.
- 7. All hazardous materials should be disposed according to guidelines for hazardous waste disposal. Materials of human or animal origin should be handled as biohazardous materials and disposed of with proper precautions.
- 8. Avoid microbial contamination of reagents as it may cause incorrect results.

# Literature

- [1] J B,A P. The changing landscape of Lynch syndrome due to PMS2 mutations[J]. Clinical genetics, 2018, 94(1).
- [2] Sanne W, Brohet RM, Tops CM, et al. Lynch syndrome caused by germline PMS2 mutations: delineating the cancer risk. J Clin Oncol. 2014, 33:319-325.
- [3] Clendenning M, Hampel H, LaJeunesse J, et al. Long-range PCR facilitates the identification of PMS2-specific mutations. Hum Mutat. 2006. 27: 490-495.

#### Date of publication or revision

2024-06-10

Change(s) made: -

## **Explanation of symbols**



Catalog number



Batch code



Use by



Temperature limitation



Do not use if package damaged



Consult instructions for use



Caution

PMS2 En\_Rev. 1 IM Page 2 of 2